

Certificate of Analysis

Product Name	Easy iMatrix-511
Lot No.	E19E008
Manufacturing Date	15 / May. / 2019
Expiration Date	14 / Nov. / 2020
Quality Assurance	<i>Y. Atsu</i> YUJI ATSUZAWA
Issue Date	06 / Jun. / 2019
Manufacture	Nippi, Incorporated Biological and Chemical Products Division
Address	1-1-1 Senju Midori-cho, Adachi, Tokyo 120-8601, Japan
Phone Number	+81(3)3888-5184

Test Results

Test Item	Specification	Result
Sterility Test	Negative	Negative
Integrin Binding Assay	Activity compared to the standard is 80% or more.	105%